



CERTIFICATE OF ANALYSIS

Product	EZ-10 Spin COVID-19 Viral RNA Extraction Kit
Product Code	VT9200
LOT#	O511R0X

Kit components

Components	VT9200 50 RXN
Buffer Rlysis-VG	30 ml
Universal RPE Solution	12 ml
Nuclease-free Water	5 ml
EZ-10 Spin Column	50
2 ml Collection Tube	50
Carrier RNA	1mg
Protocol	1

This is to certify that units of the above mentioned lot number were tested and found to comply with the specifications of the grade listed. Bio Basic expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. The following are the actual analytical results obtained:

Result name	Units	Specification	Test Value
Appearance		Report	Kit
Functional Test	Pass/Fail	=Pass Test	Pass test

~~Storage~~ **Storage:** RNA should be divided into conveniently sized aliquots and stored it at -20°C. Do not freeze-thaw the aliquots of Carrier RNA more than 3 times. All remaining components should be stored at room temperature. The kit is valid for 1 year at 4°C.

